

| Environmental Quality 4201 Normandy Street Bismarck, North Dakota 58503 Phone: (701) 328 -5150 Fax: (701) 328-5200

AUTHORIZATIONS

Title	Name	Signature
SOP Author	McKenzie Schick	Math SAA
Program Manager	Joshua Wert	Jahna With

QUALITY CONTROL/QUALITY ASSURANCE DOCUMENTATION

Title: Taking Field Measurements in Streams, Rivers, Lakes, and Wetlands Using Handheld Meter(s)

Type: Standard Operation Procedure #7.02 Version: 3.0 Date: 01/08/2020 Author: McKenzie Schick

REVISION HISTORY

Revision	Change Description	Date	Authorization
1	Update Department address	06/08/23	
1.1	Program Manager Update	02/06/2024	

ACKOWLEDGEMENTS

(Place to acknowledge peer reviewer)

THIS PAGE LEFT INTENTIONALLY BLANK

TABLE OF CONTENTS

1.0	SCOPE AND APPLICABILITY	.2
2.0	SUMMARY OF METHOD	.2
3.0	HEALTH AND SAFETY WARNING	.2
4.0	CAUTIONS	.3
5.0	INTERFERENCES	.3
6.0	FIELD PERSONNEL QUALIFICAITONS/RESPONSIBILITIES	.3
7.0	EQUIPMENT AND SUPPLIES	.4
8.0	PROCEDURE	.4
9.0	DATA AND RECORD MANAGEMENT	.5
10.0	QUALITY ASSURANCE AND QUALITY CONTROL	.6
11.0) REFERENCES	.6
APF	PENDIX A. Field Reporting Form	

APPENDIX B. SOP Acknowledgement and Training Form

1.0 SCOPE AND APPLICABILITY

This document presents the North Dakota Department of Environmental Quality, Division of Water Quality's (DWQ) Standard Operating Procedure (SOP) for performing field measurements using a handheld meter in rivers, streams, lakes, and wetlands. This SOP applies to all DWQ field staff, non-DWQ cooperators, and citizen volunteers. The following is summarized from the Field Guide for Collecting and Processing Stream Water Samples for the National Water Quality Assessment Program (Shelton 1994).

Measurements of pH can provide some of the most important limnological information pertaining to a water-body. The pH of a solution is a measure of the effective hydrogenion concentration. pH is in logarithmic units using a scale from 1 to 14. Water bodies with a pH of less than 7 are considered acidic, while water bodies with a pH greater than 7 are considered basic or alkaline.

Dissolved gases such as carbon dioxide, hydrogen sulfide, and ammonia appreciably affect pH. Due to this, pH must be taken in the field, as a significant change can occur within several hours or even minutes after sample collection.

Specific conductance is the reciprocal of resistance in ohms and is a measure of the capacity of water or another substance to conduct an electrical current. Specific conductance is reported in micro siemens per centimeter at 25 degrees Celsius. The specific conductance of water is determined by the types and quantities of dissolved substances in the water. Thus, specific conductance indicates the concentrations of dissolved solids in water.

The specific conductance of water may change significantly with time because of pollution, precipitation, absorption, ion exchange, oxidation, or reduction. Therefore, specific conductance should be measured in the field.

Temperature and dissolved oxygen (DO) measurements may also provide some of the most important limnological characteristics of a water body (e.g., biological and biochemical reactions). Therefore, water temperature and DO should be measured in the field.

2.0 SUMMARY OF METHODS

The handheld meter is calibrated daily, or prior to each use, using manufacturer's instructions. The calibration is logged into a calibration log (Appendix A). The probes are then lowered into the water and a measurement for temperature, DO, pH, and specific conductance is recorded.

3.0 HEALTH AND SAFETY WARNING

Field personnel should take appropriate precautions when operating watercraft and working on, in, or around water. All boats should be equipped with safety equipment such as personal flotation devices (PFD's), oars, air horn, etc. North Dakota's boating laws and rules shall be followed by all field personnel.

Field personnel should be aware that hazardous conditions potentially exist at every waterbody. If unfavorable conditions are present at the time of sampling, the sample visit should be rescheduled. If hazardous weather conditions arise during sampling, such as lightning or high winds, personnel should cease sampling and move to a safe location.

Field personnel should also be aware of wildlife, insects, and plants that could be harmful as well as heat stroke and hypothermia. A first aid kit should be accessible for any potential cuts, stings, bites, or contact with poisonous plants. Ensure there is access to water, sunscreen, insect repellant, and extra clothing.

4.0 CAUTIONS

- Watch for bubbles or wrinkles in the membrane, these will cause errors in the DO reading (membrane handheld meters).
- Calibration solutions need to be stored properly. (e.g. room temperature and not in vehicles exposed to temperature extremes).
- Use calibration checks to ensure probes are working correctly.
- Confirm guard is properly installed for every sample.
- Probes must be checked before each sampling season to ensure that they are in working condition. Send to YSI for maintenance if necessary.
- Do not use expired probes.
- Calibrate specific conductance first.
- Take surface readings in lakes prior to sample collection.
- Take readings in rivers and streams prior to sample collection or upstream of sample collection.
- If traveling and staying overnight, handheld meters should not be left in vehicles overnight and should be stored at room temperature when possible.

5.0 INTERFERENCES

<u>Membrane Handheld Meters Only:</u> The membrane of a DO electrode is permeable to additional gases other than oxygen, such as hydrogen sulfide (H_2S). Caution should be taken when using the membrane electrode in low DO waters since the presence of H_2S may lower the cell sensitivity. This interference can be reduced by frequently changing the membrane and calibrating the electrode. For optic DO probes, make sure that the sensor is clear of film and if a wiper is present, replace the wiper if dirty.

6.0 PERSONNEL QUALIFICATIONS/RESPONSIBILITIES

All personnel taking field measurements using a handheld meter must read this SOP annually and acknowledge they have done so via a signature page (see Appendix B). New field personnel must also demonstrate successful performance of the method. The signature page will be signed by both trainee and trainer to confirm that training was successfully completed and that the new personnel is competent in carrying out this SOP. The signature page will be kept on-file at DWQ along with the official hard copy of this SOP.

7.0 EQUIPMENT AND SUPPLIES

- Handheld meter(s)
- □ Maintenance kit (Potassium chloride/KCI solution (membrane meters only), spare membranes (membrane meters only), batteries, battery charger)
- □ Site Map
- Bathymetric lake map for lake sampling, if available
- □ Field report form
- Pens or pencils
- □ pH 7 and pH 10 calibrating buffer
- □ Specific conductance 1,413 µS cm⁻¹ calibration solution
- □ pH 9 calibration check
- □ Specific conductance 1,008 µS cm⁻¹ calibration check
- □ Power ice auger (winter sampling)
- □ Ice skimmer (winter sampling)
- Meter stick (winter sampling)
- Sled (winter sampling)
 Personal Flotation Device
- □ Boat or canoe. if needed

8.0 PROCEDURE

Streams and Rivers

- Calibrate the meter using a solution with known specific conductance (1,413 µS cm⁻ 1. ¹ preferred), pH buffer solutions of 7.0 and 10.0, and calibrate dissolved oxygen according to barometric pressure following manufacturer's instructions.
- 2. Confirm calibration effectiveness using calibration check solutions.
- 3. Record calibration information in the equipment calibration log (Appendix A). Note: Calibration logs are meter specific.
- 4. Locate the main current of the stream or river. Note: When drilling a hole through the ice, be sure not to disturb the sediment with undue agitation.
- Place the guard over the exposed probes and lower the probe to that depth which 5. is approximately 60% the total water depth below the surface. For example, if the stream is five feet deep, take the measurement three feet below the surface.
- Read temperature, dissolved oxygen, pH, and specific conductance if using a multi-6. probe meter and record. If using a single probe meter wait for the temperature reading to stabilize (30 seconds minimum), record the temperature reading on the stream and river sampling field log (Appendix A), switch the display to read dissolved oxygen, allow the dissolved oxygen reading to stabilize and record the dissolved oxygen concentration on the field report form. NOTE: To achieve an

accurate reading some units require a stirring unit or for the sampler to gently move the probe up and down two to three inches to circulate water across the membrane (membrane handheld meters only).

Lakes and Wetlands

- 1. Locate the desired sampling location and anchor boat or drill a hole through ice. Note: When drilling a hole through the ice do not disturb the water column with undue agitation.
- 2. Calibrate to manufacturer's instructions and record calibration information in the equipment calibration log (meter specific).
- 3. Fill in the station identification information on the field report form. Also, measure and record ice thickness and snow depth in the comments section (winter sampling) (Appendix A).
- 4. Remove the storage cup and replace it with a protective guard. Lower the probe to 0.5 meters depth, or just below the ice.
- 5. Read the temperature, DO, specific conductance, and pH concentrations and record on lake or wetland sample record form (Appendix A). If using a single probe meter wait for the temperature reading to stabilize (30 seconds minimum), record the temperature reading on the field report form, switch the display to read dissolved oxygen, allow the dissolved oxygen reading to stabilize and record the dissolved oxygen concentration on the field report form. NOTE: To achieve an accurate reading some units require a stirring unit or for the sampler to gently move the probe up and down two to three inches to circulate water across the membrane (membrane handheld meters only).
- 6. Lower the probe to the next depth interval and repeat step 5. Readings should be taken at every meter (i.e. 0.5, 1, 2, 3...) if greater than three meters deep. Every half meter (i.e. 0.5, 1, 1.5, 2, 2.5...) if three meters depth or less.
- 7. Repeat step 6 until 0.5 meters of the bottom.
- 8. Retrieve probe from bottom of water body, rinse thoroughly and replace the storage cup. Recheck the surface reading to ensure the reading is within 0.2 mg L⁻¹ of initial reading, following profile measurements to ensure the precision and accuracy of the measurements. If the reading is not within 0.2 mg L⁻¹, re-calibrate the meter and measure the profile again, repeating steps 5 through 7.

9.0 DATA AND RECORDS MANAGEMENT

Handheld meter measurements will be recorded on the field form (Appendix A). Once personnel reach the office, data recorded on the field form are entered into the DWQ Sample Identification Database (SID). Field notes should be used to record any quality control activity performed such as measurements taken by more than one sampler, or to record any sampling conditions that may have interfered with the reading such as high

winds/wave action, cattle in water, observed flow, water surface, water clarity, water color, water odor, visual algae cover, number of dead fish, present weather, estimated inches of rain fall in past 72 hours, and any comments. Field forms and notes should be stored in the appropriate project folder at DWQ.

10.0 QUALITY ASSURANCE AND QUALITY CONTROL

The meter(s) should be calibrated before sampling trip following the manufacturer's instructions and the calibrations should be recorded. Additionally, temperature probes will be checked annually against a certified NIST thermometer.

11.0 ANS DECONTAMINATION

In waters that have been classified as ANS infested, the meter must me decontaminated (Decon). To decon the meter pH 4.00 solution will be used. Triple rinse probes using pH 4.00 solution and then store the probes in pH 4.00 solution.

11.0 REFERENCES

Gibs, Jacob, et al. "Use of Multiparameter Instruments for Routine Field Measurements." *USGS*, USGS, Mar. 2012, https://pubs.usgs.gov/twri/twri9a6/twri9a68/twri9a6_6.8.pdf.

Shelton, Larry R. "Field Guide for Collecting and Processing Stream-Water Samples for the National Water-Quality Assessment Program." *USGS*, USGS, 1994, https://pubs.usgs.gov/of/1994/0455/report.pdf.

APPENDIX A

Field Reporting Forms

River and Stream Sampling Field Log North Dakota Department of Environmental Quality Division of Water Quality - Watershed Management Program Telephone: 701-328-6140 Fax: 701-328-6280

Sample	e #:	Site ID:	Site Description:		Comments:
Dup	Blk	Date: / /	Temperature	DO	
		Time: :	SC	рН	
Sample	e #:	Site ID:	Site Description:		Comments:
Dup	Blk	Date: / /	Temperature	DO	
		Time: :	SC	рН	
Sample	e #:	Site ID:	Site Description:	•	Comments:
Dup	Blk	Date: //	Temperature	DO	
		Time: :	SC	рН	
Sample	e #:	Site ID:	Site Description:		Comments:
Dup	Blk	Date: / /	Temperature	DO	
			SC	рН	
Sample	e #:	Site ID:	Site Description:		Comments:
Dup	Blk	Date://	Temperature	DO	
		Time: :	SC	рН	
Sample	e #:	Site ID:	Site Description:		Comments:
Dup	Blk	Date://	Temperature	DO	
		Time: :	SC	рН	
Sample	e #:	Site ID:	Site Description:		Comments:
Dup	Blk	Date: / /	Temperature	DO	
		Time: :	SC	рН	
Sample	e #:	Site ID:	Site Description:		Comments:
Dup	Blk	Date://	Temperature	DO	
		 Time: :	SC	рН	



Lake Profile Field LogNorth Dakota Department of Environmental QualityDivision of Water Quality – Watershed Management ProgramTelephone:701-328-6140Fax:701-328-6280

Project Code:					F	Project Name:				
Site Identification	1:				Sampler(s):					
Site Description:										
Date: /	/		Time:	:	A	Ambient Temp:		Wind Speed:	(mph)	
Wind Direction:			% Cloud	Cover:	S	Secchi Disk:	(m)	Baro:	(mm/Hg)	
Chlorophyll-a:			Phytopla	nkton:	I	nitial DO:		Final DO:		
Sample Depths: _	Me	eters,		_Meters,		Meters,	Me	ters		
Comments:										
Depth (m)	Temp (c)	DO (M	(Ig/L)	pН		Specific Conduct	Comments			
						Conduct				

Depth (m)	Temp (c)	DO (Mg/L)	рН	Specific Conduct.	Comments



Wadable Wetland Field Sampling FormNorth Dakota Department of Environmental QualityDivision of Water Quality – Watershed Management ProgramTelephone:701-328-6140Fax:701-328-6280

Site ID:	Wetl	and Na	Name or description:					County:		
Observer:	Observer: Date: Time:									
Aquatic Zone De	escription:						I			
Ambient Temp:			Wind Speed:				Wind D	Direction:		
Cloud Cover	% of 100	Secch	i (m)	chl	orop	ohyll-a=(y / n)		Phytoplank	xton=(y / n)	
Water Chemistry	Water Chemistry Taken At		Meters,			Meters,		Meters,		Meters,
Depth (m) or Location	Tem	p. (C)	D.O. (mg/L)	pH	I	Conductivity (umhos/cm)	Comm	ients		
Site Drawing:										

APPENDIX B

SOP Acknowledgement and Training Form

SOP Acknowledgement and Training Form

This SOP must be read and this form signed annually. This form must be kept with the latest version of the SOP.

Document Title:	
Document Revision Number:	
Document Revision Date:	

Please sign below in accordance with the following statement:

"I have read and understand the above referenced document. I agree to perform the procedures described in this SOP in accordance with the document until such time that it is superseded by a more recent approved revision."

Printed Name	Signature	Date

SOP Acknowledgement and Training Form (con't)

<u>Trainee</u>: Sign below to acknowledge that training on this SOP was received, understood, and all questions/concerns were addressed by the trainer.

<u>Trainer</u>: Sign below to acknowledge that training on this SOP was completed for the individual listed and that training is competent to perform the procedures described within.

Date of Training	Trainee Printed Name	Trainee Signature	Trainer Printed Name	Trainer Signature